



Product Information

Glove Name Curad® Latex Gloves
Description Powder-Free Latex Examination Gloves
Non-Sterile, Ambidextrous

US Office: Medline Industries, Inc.
Northfield, IL, USA
1-800-MEDLINE
www.medline.com

Product Numbers		Packaging
CUR8103	X-Small	100 Each/Box
CUR8104	Small	10 Boxes/Case
CUR8105	Medium	1000 Each/Case
CUR8106	Large	(90 ea/bx, 10 bx/cs for XL)
CUR8107	X-Large	

Physical Properties and Quality Characteristics

Characteristics:	Glove Color/Finish	Beige/Fully Textured
	Glove Length	240 mm
	Finger Thickness	0.12 mm (4.7 mils)
	Palm Thickness	0.11 mm (4.3 mils)
	Primary Material	Latex
	Accelerators	Carbamates
	Cuff	Beaded

Pinholes:	Acceptable Quality Level (AQL)	1.5
	FDA Requirement	2.5

Tensile Strength:	Before Aging/After Aging	23MPa/18MPa
	ASTM Standard D3578	18 MPa/14 MPa

Ultimate Elongation:	Before Aging/After Aging	740%/660%
	ASTM Standard D3578	650%/500%

Viral Penetration Test:	ASTM Standard F1671	Complies with F1671 Test Method
--------------------------------	---------------------	---------------------------------

Natural Rubber Latex Protein Content:	ASTM Standard D5712	Less than 50 micrograms per gram of glove
--	---------------------	---

Residual Powder Content:	ASTM Standard D6124	Less than 2 mg/glove
---------------------------------	---------------------	----------------------

CAUTION: This product contains natural rubber latex which may cause allergic reactions.

Curad and Medline are registered trademarks of Medline Industries, Inc.

© Medline Industries, Inc. not to be reproduced or distributed without prior written permission of Medline Industries, Inc.

Updated 6.2021